

CDC 2019-nCoV Real-Time RT-PCR Diagnostic Panel

Acceptable Alternative Primer and Probe Sets

The following lots of N1, N2, and RP primers and probes have passed functional testing at CDC and may be used with the CDC 2019-nCoV Real-Time RT-PCR Diagnostic Panel Instructions for Use under CDC's Emergency Use Authorization (EUA):

<https://www.fda.gov/medical-devices/emergency-situations-medical-devices/emergency-use-authorizations>

Please note that only the listed lot numbers of material are acceptable for testing under CDC's EUA.

Any primer and probe reagents included in these kits in addition to N1, N2 and RP have not been tested by CDC and may not be used for diagnostic testing under CDC's EUA. Only N1, N2 and RP primer and probe sets may be used.

Manufacturer	Product name	Catalog No.	Acceptable Lots
BioSearch Technologies www.biosearchtech.com 1-800-GENOME-1	2019-nCoV CDC Probe and Primer Kit for SARS-CoV-2	KIT-nCoV-PP1- 1000	143503 143764 144000
Integrated DNA Technologies (IDT) www.idtdna.com (800) 328-2661	2019-nCoV Kit, 500 rxn	10006606	0000500383 0000504847 0000505594 0000505969 0000507295 0000507509 0000508150 0000508785 0000509402 0000510344 0000510908

Last Updated: March 30, 2020

For current list please visit: <https://www.cdc.gov/coronavirus/2019-ncov/lab/index.html>